

# Calibration of Thermometers (Lecture and Training)

8 Nov. 2016

***Hiroshi YOSHIMURA***

***Scientific Officer***


***Regional Instrument Centre Tsukuba***

***Observing Division, Observing Department***

***Japan Meteorological Agency***

# Outline


1. [Theory] Temperature measurements and calibration of thermometer  
(Here in the lecture room)
  
2. [Practice] Calibration of thermometers  
(At the calibration room on the 1<sup>st</sup> floor)




# 1. Temperature measurements (theory)

# Definition of the SI unit of thermodynamic temperature (kelvin)

"The kelvin, unit of thermodynamic temperature, is the fraction 1/273.16 of the thermodynamic temperature of the triple point of water."



Water triple point cell



# The International Temperature Scale of 1990 (ITS-90)

➤  $t/^\circ\text{C} = T/\text{K} - 273.15$

T: thermodynamic temperature (unit: Kelvin)

t: temperature in degrees Celsius (unit:  $^\circ\text{C}$ )

- The range from 13.8033K(-259.3467 $^\circ\text{C}$ ) to 961.78 $^\circ\text{C}$  is defined by means of platinum resistance thermometers calibrated at certain fixed points and applying specified interpolation procedures.

# Defining fixed points of the ITS-90

Number	Temperature		Substance(*1)	State(*2)
	T90/K	t90/		
1	3 to 5	-270.15 to -268.15	He	V
2	13.8033	-259.3467	e-H <sub>2</sub>	T
3	~ 17	~ -276.15	e-H <sub>2</sub> (or He)	V(or G)
4	~ 20.3	~ -252.85	e-H <sub>2</sub> (or He)	V(or G)
5	24.5561	-248.5939	Ne	T
6	54.3584	-218.7916	O <sub>2</sub>	T
7	83.8058	-189.3442	Ar	T
8	234.3156	-38.8344	Hg	T
9	273.16	0.01	H <sub>2</sub> O	T
10	302.9146	29.7646	Ga	M
11	429.7485	156.5985	In	F
12	505.078	231.928	Sn	F
13	692.677	419.527	Zn	F
14	933.473	660.323	Al	F
15	1234.93	961.78	Ag	F
16	1337.33	1064.18	Au	F
17	1357.77	1084.62	Cu	F

(\*1)All substances except 3He are of natural isotopic composition;

e-H<sub>2</sub> is hydrogen at the equilibrium concentration of the ortho- and para-molecular forms.

(\*2)V: vapour pressure point;

T: triple point (temperature at which the solid, liquid, and vapour phases are in equilibrium);

G: gas thermometer point;

M ,F: melting point, freezing point

(temperature, at a pressure of 101325Pa, at which the solid and liquid phases are in equilibrium)





# Types of thermometers

## 1. Contact-type thermometers

- Platinum resistance thermometer
- Liquid-in-glass thermometer
- Thermocouple
- etc.

## 2. Radiation thermometers


# Platinum resistance thermometer



Example of the relation between temperature and resistance


# Traceability of Temperature (JMA)

Jan. 2013 by JMA



# Traceability of Temperature (JMA)

Jan. 2013 by JMA





**JCSS**

SERIAL No. 120674A

**CERTIFICATE OF CALIBRATION**

ACCREDITED LABORATORY  
Tanaka Kikinzoku Kogyo K.K. Isehara Works Thermometer Calibration Laboratory

ADDRESS Suzukawa 26 Isehara-City Kanagawa Japan 259-1146

Meteorological  
Customer : Instruments Center JMA Address: 1-2 nagamime, tsukuba ibaraki

Item : Platinum Resistance Thermometer Model No.: P-1000

Type : NSR-160 Serial No.: NSR-160

Calibration Method: Fixed Point Ambient : Temp. 25°C, Hum. 60%

Description : Temperature

**COPY**

The Calibration Result.

Temperature (°C)	Resistance Ratio(Wt)
Triple Point of Mercury -38.8344	0.844216 ±0.000008
Freezing Point of Indium 156.5985	1.609468 ±0.000012


The uncertainties are for a confidence probability of not less than 95%

Note : All quoted resistances ratio has been extrapolated to a current of 0mA.  
The resistance at triple point of water has been 99.9289Ω.  
The measurement has done by using Guildline 9975 current comparator resistance bridge.

Calibrated on 25 August 2012

Issued by Tanaka Kikinzoku Kogyo K.K.  
Isehara Works Thermometer Calibration Laboratory

Date of issue 28 August 2012

Approved Signature   
Tokio HAMADA

This certificate is issued in accordance with the conditions of accreditation granted by Japan Calibration Service System, which is based on ISO/IEC 17025 and also participates in the MRA not only APLAC but also ILAC.

# Certificate of Calibration (Platinum resistance thermometer)


## The Calibration Result.

Temperature (°C)	Resistance Ratio(Wt)
Triple Point of Mercury -38.8344	0.844216 ±0.000008
Freezing Point of Indium 156.5985	1.609468 ±0.000012




# Traceability of Temperature (JMA)


Jan. 2013 by JMA



# Meteorological Instrument Center

Japan Meteorological Agency







**JCSS**  
JCSS 0295

## Calibration Certificate

Client Name: \*\*\*\*\*

Client Address: \*\*\*\*\*

Calibration Site: Meteorological Instruments Center, Japan Meteorological Agency

Calibration Item: Thermometer

Type and Serial Number: \*\*\*\*\*

Manufacturer: \*\*\*\*\*

Calibration Method: As shown in page 2.

Calibration Conditions: Room Temperature 21.6 °C - 25.6 °C.

Relative humidity 22 % - 34 %.

Calibration Results: As shown in page 2.

Date of Application: 26 November 2012.

Date of Performance of the Calibration: 27 - 29 November 2012.

Date of issue: 10 January 2013.

The issuing Authority:

+

+

Mitsugi Katsuyama

Director,

Meteorological Instruments Center, JMA

1-2 Nagamine Tsukuba-shi Ibaraki Japan

This certificate is based on article 144 of the Measurement Law and indicates the result of calibration in accordance with measurement standards traceable to Primary Measurement Standards (National Standards) which realizes the physical units of measurement according to the International System of Units (SI). The accreditation symbol is attestation of which the result of calibration is traceable to Primary Measurement Standards (National Standards).

The certificate shall not be reproduced except in full, without the written approval of the issuing laboratory.

The calibration laboratory who issued this calibration certificate conforms to ISO/IEC 17025:2005.

This calibration certificate was issued by the calibration laboratory accredited by IAJapan who is a signatory to the Mutual Recognition Arrangement (MRA) of International Laboratory Accreditation Cooperation (ILAC) and Asia Pacific Laboratory Accreditation Cooperation (APLAC). This (These) calibration result(s) may be accepted internationally through ILAC/APLAC MRA.

### Calibration method

- The calibration item was calibrated with the following reference standards and instruments:

Platinum resistance thermometer: NSR-160, NS-05030 (Netsushin)

Water triple point cell: 5901C-G, 2038 (FLUKE)

Alternating current bridge: F600AC, 012705/09 (ASL)

Standard resistor 100 Ω in temperature controlled enclosure:

5685A, 14553/02 (Tinsley)

- The calibration item was compared with the above reference standards in the liquid chamber at the following calibration points other than 0 °C (in the ice point bath). The procedure used in this calibration was the standard operating procedure manual No.4 of Meteorological Instruments Center, JMA

### Calibration results

Nominal temperature (°C)	Reference temperature (A) (°C)	Calibrated indication temperature (B) (°C)	Deviation (B) - (A) (°C)	Expanded uncertainty (°C)
-10	-10.000	-9.980	0.020	0.045
0	0.000	0.020	0.020	0.013
10	10.000	10.019	0.019	0.036
20	20.000	20.019	0.019	0.036
30	30.000	30.018	0.018	0.036
40	40.000	40.019	0.019	0.036
50	50.000	50.020	0.020	0.036


### Note

The reported expanded uncertainty is stated as the combined standard uncertainty multiplied by the coverage factor  $k = 2$ , which for a normal distribution corresponds to a coverage probability of approximately 95 %.

Characterization type is ITS-90 ( $R_0 = 99.9963$ ,  $A_4 = -1.938829 \times 10^4$ ,

$B_4 = 1.068843 \times 10^4$ )


- End of Certificate -



# Calibration result


## Image

Distribution of Calibrated  
Indication of temperature



# Traceability of Temperature (JMA)

Jan. 2013 by JMA






<Not 0°C>

Alternating current  
bridge  
(Working standard)


Platinum resistance  
thermometer  
(Working standard)



<0°C>



Platinum resistance  
thermometer  
or  
Mercury-in-glass  
thermometer  
(Field instruments)



Liquid bath chamber




Ice point  
inspection box

# An example of Calibration result of Mercury-in-glass thermometer


Instrument No.	測器識別番号	
Calibration point	検査点 [°C]	補正值 [°C]
	41625-00215	
	40	-0.1
	35	-0.1
	30	-0.2
	25	-0.2
	20	-0.1
	15	-0.1
	10	-0.1
	5	-0.1
	0	-0.1
	-5	-0.1
	-10	-0.1
	-15	-0.2
	-20	-0.2
	-25	-0.2
	-30	-0.2

Instrument error

# Traceability of Temperature as to JMA



# An example chart of traceability (Temperature)



## 2. Calibration of thermometer (practice)


# Today's practice (planned...)

➤ Calibrated items:


- Platinum resistance thermometer: 1 sets
- Mercury-in-glass thermometer : 3 sets

➤ Calibration points:

- 0°C (Ice point)
- 30°C (with liquid bath chamber)




# Liquid in glass thermometer



**Caution for reading:**

The observer ensures  
that the straight line  
from his/her eye to the  
meniscus, or index.

# Liquid in glass thermometer



How to read the value;

- The interval of line:  $0.2^{\circ}\text{C}$
- Read the value in units of  $0.1^{\circ}\text{C}$ .

The value of left example;

$29.7^{\circ}\text{C}$




# 3. Minamata Convention on Mercury

# Minamata Convention on Mercury

- The Minamata Convention on Mercury is a global treaty to protect human health and the environment from the adverse effects of mercury, which was agreed in 2013.
- Under this Convention, imports and exports of mercury will no longer be allowed. In this context, production, importation and exportation of mercury-added products such as thermometers and barometers will be stopped by the year 2020.




NMHSs which use mercury-added instruments should prepare a transition to modern Alternatives.



<http://www.mercuryconvention.org/>



**Thank You!**



Mascot of JMA "Harerun"

