


Meteorological Instrument
Thai Meteorological Department
THAILAND

by

MR PISOOD PROMSUT



WMO/JMA Training Workshop
On
Calibration and Maintenance of Meteorological
Instrument in RAII at TSUKUBA
JAPAN



METEOROLOGICAL INSTRUMENT BUREAU
THAI METEOROLOGICAL DEPARTMENT

4353 meteorological department sukhumvit rd. bangna
bangkok thailand 10260

บทนำ

- เนื่องจากเครื่องมือวัดที่ใช้ในการตรวจวัดสภาพอากาศมีผลกระทบต่อความแม่นยำและเชื่อถือได้ของข้อมูลจึงจำเป็นต้องดำเนินการให้แน่ใจว่าเครื่องมือวัดที่ใช้ในการตรวจวัดสภาพอากาศสามารถให้ผลการวัดที่แม่นย้ำตามความต้องการ การสอบเทียบ (Calibration) เป็นกิจกรรมที่จำเป็นสักครั้ง การทำให้แน่ใจว่าเครื่องมือวัดที่ใช้งานยังสามารถทำงานได้อย่างแม่นย้ำตามที่ต้องการ



METEOROLOGICAL DEPARTMENT OF THAILAND

Meteorological Instrument

There are many type of meteorological instrument

- 1 Weather satellite system
- 2 Weather radar system
- 3 Automatic weather station more than 90 station
- 4 Surface weather station more than 100 station
- 5 Automatic Rain Gauge Station
- 6 Seismic Monitoring Station 40 Station



3

5

4



ความหมายของการสอบเทียบ

- การสอบเทียบ หมายถึง เป็นชุดค่าเนินการภายในตัวสภาวะเฉพาะเพื่อหาค่าความสัมพันธ์ระหว่าง เครื่องมือวัดเพื่อเปรียบเทียบค่าที่รู้ของปริมาณที่วัด (ซึ่งต้องเป็นค่าที่สามารถอ้างอิงได้) และ จากการสอบเทียบจะให้ข้อมูลว่าเครื่องมือวัดที่ใช้ในการตรวจสอบสภาพอากาศยังคงมีคุณลักษณะ ทางด้านมาตรฐานที่เหมาะสมในการใช้งานต่อไปหรือไม่



AGROMETEOROLOGICAL STATION



Surface Meteorological Station



Maximum Thermometer
Minimum Thermometer
Psychrometers exposure
(wet bulb, dry bulb)



Mercury Barometer



Rain Gauge



wind speed and wind direction

Inspection System of Thai Meteorological Department

Meteorological Standard

Thai Meteorological
Department (TMD)

Japan
Meteorological
Agency (JMA)

2 years

Users
Laboratories, Factories,
Manufacturers



WMO REGIONAL INSTRUMENT CENTRES (RICS)

Regional Associations	Region	Instrument Centres
RA I	Africa	Algiers (Algeria) Cairo (Egypt) Casablanca (Morocco) Nairobi (Kenya) Gaborone (Botswana)
RA II	Asia	Tsukuba (Japan) Beijing (China)
RA III	South America	Buenos Aires (Argentina)
RA IV	North and Central America	Bridgetown (Barbados) San Jose (Costa Rica) Mount Washington (USA)
RA V	South. West Pacific	Manila (Philippines) Melbourne (Australia)
RA VI	Europe	Bratislava (Slovakia) Ljubljana (Slovenia) Trappes (France)



2-4 April 1979 comparision at Kolkata INDIA

4-6 Jun 1984 comrarision at Tokyo Japan

27 Feb – 11 March 1994 Comparision at
Tokyo Japan

2007Comparision at RIC (TSUKUBA) Japan

2010 comparision at RIC (TSUKUBA) Japan

National Standard Fortin Barometer

Index Error Pressure : + 0.03 mm.Hg

Temp : - 0.16 ° C



Calibration barometer and thermometer At RIC (Regional Instrument Centre) in tsukuba Japan : Year 2007



Fortin barometer



Sub Standard Travelling
kew mercury barometer
Serial No 610 /440228
Type 00.06100.100000
No. 702055.006
Made in Germany



Digital barometer



Digital barometer
Model PTB220ACB2A3
Serial V1220015
Vaisala Finland



Glass thermometer
(Normal) Serial 6169



Wind tunnel
type 8420



Thermal anemometer f642



Ultrasonic
anemometer



Pitot tube



Micromanometer fc 014

Test chamber for Pressure

Calibration



National standard baro

Sub standard barometer
Manufacturer negretti & ZAMBRA LONDON
Serial No. M/7620
Range : 50 – 850 mm.Hg.

Vacuum chamber

Manufacturer : Negretti & ZAMBRA

Made in england

Inside dimensions 20.5 x 17 x 13 inch

For test meteorologigal Instrument,

- Aneroid barometer
- Digital barometer
- Barograph
- Micro-barograph





Mercury
barometer



Sgreen Chamber



Psychrometer (dry-
Wet)



RAIN
Gauges



Rainfull recorder



Grass minimum thermo

SUNSHINE
RECORDERS




Meteorological
Instrument



Calibration for barometer



A plan of Practical use of TMD's Traveling Standard instrument



Mercury barometer
Lambrecht
no.610/... in
Germany

Digital
barometer
PTB220(Vaisala ,Finland)


Glass
Thermometer

Asmann
psychrometer

Environment test chamber For temperature and humidity

Spec

-temperature
range -40 – 200 ° C
Model no T10RS
Serial no 26042-09
USA.



Grass thermometer
Digital thermometer

Thermograph

Standard thermometer Serial No. 918802



Temp. At Chamber Adjusted approximate
We use a Glass Thermometer as a temperature
Standard

Adjust the temperature seting between 2 to 45
degree

Problem

- The cooling time is very slow(very time consuming)
- If we raise the temperature to be faster (OK.)
- When we turn on to machines working to be very noisy
- This is Instrument Long a lifetime



Humidity Test Cabinet

Technic Data

Max error in measurement of Psychrometer $\pm 1\%$

Scale division of psychrometer 0.2 degree
Accuracy of hygrometer for humidity control $\pm 2.5\%$ relative humidity, if the measuring element is regenerated at regular intervals

Test chamber H x W x D
400 x 700 x 280 mm.



Weight approx.
170 kg



Hygrometer

Aspirated Psychrometer



Humidity Test Cabinet

The calibration test chamber is designed for the calibration of humidity sensors. It is possible to calibrate several sensors at the same time. The operating range is from 15% to 95% rel. humidity

Technical specification

Operating range

Power supply

Volume test chamber
weight

Variations

<15 ... 96 %rel. humidity ± 1 %

230 / 50 Hz

Approx. 80 l.

Approx. 170 kg.



Supply voltage 220 v AC / 50 Hz
Max. power supply 3 kva
Usable test chamber volume approx 80 l.
Humidity control approx. 15 - 95 % Rh.
Use Silica Gel powder is Hygroscopic (adsorption)

Type 8222.0000 MADE IN GERMANY

For Aspirated Psychrometer By ASSmann
Scale divisions 0.2 ° C
Size length 440 mm
weight approx. 1.5 kg

Manufactures THEODOR FRIEDRICH & CO LTD.
MADE IN GERMANY



Process of wind speed calibration



Ultrasonic
Anemometer
 $0.5 \rightarrow 20 \text{ m/sec}$

Air Pressure Measuring
Instruments
 $1 \rightarrow 30 \text{ m/sec}$

Standard Wind Speed
 $0.5 \rightarrow 20 \text{ m/sec}$

Standard Wind Speed
 $20 \rightarrow 30 \text{ m/sec}$

Calibration for wind speed

Users

Factories

Manufactures

Laboratories

Ultrasonic
Anemometer Kaijo
Sonic Corporation
DA-650-3TV
(TR-90AH)

Japan Quality
Assurance
Organization (JQA)

5 years

6 months

Pressure Measuring
Instruments
Dwyer No. 1425 Hook
Gage

Department of
Science Service
Ministry of
Science and
Technology

5 years

National Standard Wind Tunnel



Design : Open-Circuit Wind Tunnel
Measuring range : 0.15 ... Approx. 50 m/sec
Dimensions : 652 cm.



ULTRASONIC ANEMOMETER Kaijo Sonic Corporation

MODEL DA-650 3TV (Sensor TR-90AH)
Serial Number 110730029 (Sensor
120629586)

Sensor Box Size : 39 x 39 x 20 cm.
Display : 48 x 35 x 15 cm.

Measurement range : Standard type
0 m/s to 30 m/s (TR-90AH Type
probe)

Accuracy of operation : Within 1 %
Resolution : 0.005 m/s or



Windtunnel Specification	Variations
Measuring range	0.15.....50 m/s
According reynold number	0.65×10^42.15×10^4
	(at $p_l = 1013$ hPa and $t_l = 20^\circ C$)
Degree of turbulence (measured in the tunnel)	at $v = 3$ m/s $\pm 0.4\%$ at $v = 10$ m/s $\pm 0.4\%$
Cross section with :	at $v = 40$ m/s $\pm 0.3\%$
Drive power	43 kW
Drive control	SCR
Power supply	230 / 400 v 50 Hz / 3 Ph max. 50 kVa

Theodor Friedrichs & Co. MADE IN GERMANY









