


Burettes (JMA working standard)



『1,571ml burette(50mm precipitation)』

Inspected in every 10 years to the upper working standard at local government calibration centers.



- The volume of burette:1571ml
(equivalent to 50mm precipitation)
 $1571\text{ml} = 50\text{mm} \times 10\text{cm} \times 10\text{cm} \times$
(diameter of tipping bucket gauge :20cm)
- Tipping bucket gauges are inspected at 80mm/h,
20mm/h of precipitation intensity